

Certificate of Compatibility

Auckland Medical Polymer (Tianjin) Co., Ltd, certifies that the following products, which are manufactured according to the requirements of a class I product, and which meet the provisions of the directive 93/42/EEC, and the registered Quality System Registration BSI ISO 13485:2016 (Certificate No. MD684799):

AHD001	Okland [®] Patient Return Electrode (PRE-0830)
AHD003	Okland [®] Pediatric Patient Return Electrode (PRE-0840)
AHD004	Okland [®] Dual Cord Patient Return Electrode (PRE-0835)
AHD006	Okland [®] Patient Return Electrode – (PRE-0845)

when used in combination with Okland[®]

CC-01	Connecting Cable (Single pole), 4.1 m OR
CC-02	Connecting Cable (Dual Pole), 4.1 m

are compatible with the electrosurgical generators listed below:

Manufacturer: **Valleylab™**

Model/Type: **Force Triad™**
Force FX
Force 2
FT10
FX 8

Tianjin, January 2020

Jessie Gao
Vice-President
Auckland Medical Polymer (Tianjin) Co., Ltd.

Certificate of Compatibility

Auckland Medical Polymer (Tianjin) Co., Ltd, certifies that the following products, which are manufactured according to the requirements of a class I product, and which meet the provisions of the directive 93/42/EEC, and the registered Quality System Registration BSI ISO 13485:2016 (Certificate No. MD684799):

AHD001	Okland [®] Patient Return Electrode (PRE-0830)
AHD003	Okland [®] Pediatric Patient Return Electrode (PRE-0840)
AHD004	Okland [®] Dual Cord Patient Return Electrode (PRE-0835)
AHD006	Okland [®] Patient Return Electrode – (PRE-0845)

when used in combination with Okland[®]

CC-01 Connecting Cable (Single pole), 4.1 m

are compatible with the electrosurgical generators listed below:

Manufacturer: **ERBE**

Model/Type: **ICC350**
 ICC300
 ICC200

Tianjin, January 2019

Jessie Gao
Manager
Auckland Medical Polymer (Tianjin) Co., Ltd.